




ADAPTER AutoISO-2511



INSTRUKCJA OBSŁUGI

Wersja 1.00 20.01.2023



Adapter AutoISO-2511 – płyta czołowa



- Adapter AutoISO-2511 nie jest samodzielnym przyrządem pomiarowym. Jest akcesorium przystosowanym do współpracy z dedykowanym miernikiem (**MIC-2511**). Zasady używania adaptera i wykonywania nim pomiarów znajdują się w instrukcji miernika. Przed przystąpieniem do pomiarów prosimy się z nią zapoznać.
- Producent zastrzega sobie prawo wprowadzania zmian w wyglądzie, wyposażeniu i danych technicznych adaptera.

1 Informacje ogólne

1.1 Symbole bezpieczeństwa

Poniższe symbole zostały użyte na przyrządzie i/lub w niniejszej instrukcji:

	Ostrzeżenie. Zobacz wyjaśnienie w instrukcji obsługi
	Uziemienie
	Podwójna izolacja (klasa ochronności)
	Deklaracja zgodności z dyrektywami Unii Europejskiej (Conformité Européenne)
	Nie wyrzucać z innymi odpadami komunalnymi
	Uwaga, ryzyko porażenia prądem elektrycznym. Przyrząd generuje napięcie na poziomie 2500 V

Kategorie pomiarowe według normy PN-EN IEC 61010-2-030:

- **CAT II** – dotyczy pomiarów wykonywanych w obwodach bezpośrednio dołączonych do instalacji niskiego napięcia,
- **CAT III** – dotyczy pomiarów wykonywanych w instalacjach budynków,
- **CAT IV** – dotyczy pomiarów wykonywanych przy źródle instalacji niskiego napięcia.

1.2 Bezpieczeństwo

Adapter AutoISO-2511 służy do pomiarów rezystancji izolacji kabli i przewodów 3-, 4-, i 5-żyłowych we współpracy z miernikiem nadrzędnym. Aby zapewnić odpowiednią obsługę należy przestrzegać następujących zaleceń:

- Przed rozpoczęciem eksploatacji przyrządu należy dokładnie zapoznać się z niniejszą instrukcją i zastosować się do przepisów bezpieczeństwa i zaleceń producenta.
- Każde inne zastosowanie przyrządu niż podane w tej instrukcji może spowodować jego uszkodzenie i być źródłem poważnego niebezpieczeństwa dla użytkownika.
- Adapter powinien być obsługiwany wyłącznie przez osoby odpowiednio wykwalifikowane posiadające wymagane uprawnienia do przeprowadzania pomiarów w instalacjach elektrycznych. Posługiwanie się adapterem przez osoby nieuprawnione może spowodować jego uszkodzenie i być źródłem poważnego niebezpieczeństwa dla użytkownika.
- Adaptera nie wolno stosować do sieci i urządzeń w pomieszczeniach o specjalnych warunkach, np. o atmosferze niebezpiecznej pod względem wybuchowym i pożarowym.
- Niedopuszczalne jest używanie:
 - ⇒ przyrządu, który uległ uszkodzeniu i jest całkowicie lub częściowo niesprawny,
 - ⇒ przyrządu, którego przewody mają uszkodzoną izolację,

⇒ przyrządu przechowywanego zbyt długo w złych warunkach (np. zawilgoconego). **Po przeniesieniu z otoczenia zimnego do ciepłego o dużej wilgotności nie wykonywać pomiarów do czasu ogrzania przyrządu do temperatury otoczenia (ok. 30 minut).**

- W sytuacji, gdy produkt współpracuje z innymi przyrządami lub akcesoriami, stosuje się najniższą kategorię pomiarową połączonych urządzeń.
- Naprawy mogą być wykonywane wyłącznie przez autoryzowany serwis.



UWAGA!

Podanie napięcia na badany obiekt w trakcie trwania pomiaru spowoduje uszkodzenie adaptera.

2 Opis wyprowadzeń

1 DATA

Przewód do podłączenia adaptera do odpowiedniego gniazda miernika (sterowanie pracą adaptera).

2 Riso+, Riso-

Przewody do podłączenia adaptera do odpowiednich gniazd miernika z napięciem pomiarowym U_{iso} .

3 L1, L2, L3, N PEN, PE

Przewody do podłączenia adaptera do odpowiednich żył mierzonego kabla (przewodu).



UWAGA!

Podłączenie adaptera do napięcia większego niż 600 V AC może spowodować jego uszkodzenie.



Przewód do podłączenia adaptera do uziemienia.

3 Czyszczenie i konserwacja



UWAGA!

Należy stosować jedynie metody konserwacji podane przez producenta w niniejszej instrukcji.

Adapter można czyścić miękką, wilgotną flanelą używając ogólnie dostępnych detergentów. Nie należy używać żadnych rozpuszczalników, ani środków czyszczących, które mogłyby porysować obudowę (proszki, pasty itp.).

Układ elektroniczny adaptera nie wymaga konserwacji.

4 Magazynowanie

Przed złożeniem do przechowania należy upewnić się, że adapter jest suchy.

5 Rozbiórka i utylizacja

Zużyty sprzęt elektryczny i elektroniczny należy gromadzić selektywnie, tj. nie umieszczać z odpadami innego rodzaju.

Zużyty sprzęt elektroniczny należy przekazać do punktu zbiórki zgodnie z Ustawą o użytym sprzęcie elektrycznym i elektronicznym.

Przed przekazaniem sprzętu do punktu zbiórki nie należy samodzielnie demontować żadnych części z tego sprzętu.

Należy przestrzegać lokalnych przepisów dotyczących wyrzucania opakowań, zużytych baterii i akumulatorów.

6 Dane techniczne

- a) rodzaj izolacji wg PN-EN 61010-1podwójna
- b) kat. pomiarowa wg PN-EN IEC 61010-2-030IV 600 V
- c) st. ochr. obudowy wg PN-EN 60529IP40
- d) wymiary 220 x 110 x 45 mm
- e) masa (z przewodami)ok. 0,8 kg
- f) temperatura pracy-5°C...+50°C
- g) temperatura przechowywania-20°C...+60°C
- h) wilgotność względna20%...90%
- i) wilgotność odniesienia30%
- j) błąd dodatkowy ze wzgl. na wpływ wilgotności >30%-0,4% w.m. na każdy 1% ponad wilg. 30%

w.m. – wartość mierzona

7 Akcesoria standardowe

W skład standardowego kompletu dostarczanego przez producenta wchodzi:

- 4x krokodyłek czarny 11 kV 32 A – **WAKROBL32K09**,
- krokodyłek niebieski 11 kV 32 A – **WAKROBU32K09**,
- krokodyłek złoty 11 kV 32 A – **WAKROYE32K09**,
- deklaracja sprawdzenia.

Aktualne zestawienie akcesoriów znajduje się na stronie internetowej producenta.

8 Producent

Producentem przyrządu prowadzącym serwis gwarancyjny i pogwarancyjny jest:

SONEL S. A.

ul. Wokulskiego 11
58-100 Świdnica

tel. (74) 858 38 00 (Biuro Obsługi Klienta)

e-mail: bok@sonel.pl

internet: www.sonel.pl



UWAGA!

Do prowadzenia napraw serwisowych upoważniony jest jedynie producent.




AutoISO-2511 ADAPTER



USER MANUAL

Version 1.00 20.01.2023



AutoISO-2511 adapter – front panel

1 General information

1.1 Safety symbols

The following international symbols are used in the device and/or in this manual:

	Warning. See explanation in the manual
	Ground
	Double insulation (protection class)
	Declaration of Conformity with EU directives (Conformité Européenne)
	Do not dispose of with other household waste
	Attention, risk of electric shock. The device generates a voltage of 2500 V

Measurement categories according to EN IEC 61010-2-030:

- **CAT II** – concerns measurements performed in circuits directly connected to low voltage installations,
- **CAT III** – concerns measurements performed in buildings installations,
- **CAT IV** – concerns measurements performed at the source of low voltage installation.

1.2 Safety

AutoISO-2511 adapter is used to measure the insulation resistance of cables and wires with 3, 4, and 5 conductors. In order to provide conditions for correct operation, the following recommendations must be observed:

- Before you proceed to operate the device, acquaint yourself thoroughly with this manual and observe the safety regulations and specifications defined by the producer
- Any application that differs from those specified in the present manual may result in a damage to the device and constitute a source of danger for the user.
- The adapter should be operated only by suitably qualified persons having the necessary permissions to carry out measurements on electrical systems. Operating the adapter by unauthorised personnel may result in damage to the device and constitute a source of danger for the user.
- The adapter must not be used for networks and devices in areas with special conditions, e.g. fire-risk and explosive-risk areas.
- It is unacceptable to operate the device when:
 - ⇒ it is damaged and completely or partially out of order,
 - ⇒ its cable insulation is damaged,

⇒ it was stored for an excessive period of time in disadvantageous conditions (e.g. excessive humidity). **After moving the device from a cool to a warm place with a high level of relative humidity, do not start measurements until the device is warmed up to the ambient temperature (approximately 30 minutes)**

- In a situation where the product works with other instruments or accessories, the lowest measurement category of the connected devices is used.
- Repairs may be performed only by an authorized service point.



ATTENTION!
Applying voltage to the tested object during the measurement will cause damage the adapter.

2 Description of outputs

1 DATA

A cable for connecting the adapter to an appropriate port of the meter (adapter control).

2 R_{iso+}, R_{iso-}

Cables for connecting the adapter to appropriate ports of the meter with measuring voltage U_{iso}.

3 L1, L2, L3, N PEN, PE

Cables for connecting the adapter to appropriate conductors of the tested/measured cable (wire).



ATTENTION!
The adapter may be damaged when connected to the voltage exceeding 600 V AC.



A cable to connect the adapter to earth.

3 Cleaning and maintenance



ATTENTION!
Use only the maintenance methods specified by the manufacturer in this manual.

The adapter may be cleaned with a soft, damp cloth using all-purpose detergents. Do not use any solvents or cleaning agents which may scratch the casing (powders, pastes, etc.).

The electronic system of the adapter does not require maintenance.

4 Storage

Before storage, make sure that the adapter is dry.

5 Dismantling and utilization

Worn-out electric and electronic equipment should be gathered selectively, i.e. it must not be placed with waste of another kind.

Worn-out electronic equipment should be sent to a collection point in accordance with the law of waste electrical and electronic equipment.

Before the equipment is sent to a collection point, do not dismantle any elements.

Observe local regulations concerning disposal of packages, waste batteries and accumulators.

6 Technical data

- a) insulation type acc. to EN 61010-1 double
- b) measurement cat. acc. to EN IEC 61010-2-030 IV 600 V
- c) protection class acc. to EN 60529 IP40
- d) dimensions 220 x 110 x 45 mm
- e) weight (with cables) ca. 0.8 kg
- f) operating temperature -5°C...+50°C
- g) storage temperature -20°C...+60°C
- h) humidity 20%...90%
- i) reference humidity 30%
- j) additional error due to the influence of humidity >30%
.....-0.4% m.v. for every 1% above 30% humidity

m.v. – measured value

7 Standard accessories

The standard set of equipment supplied by the manufacturer includes:

- 4x crocodile clip black 11 kV 32 A – **WAKROBL32K09**
- crocodile clip blue 11 kV 32 A – **WAKROBU32K09**
- crocodile clip yellow 11 kV 32 A – **WAKROYE32K09**
- declaration of verification

The current list of accessories can be found on the manufacturer's website.

8 Manufacturer

The manufacturer of the device and provider of guarantee and post-guarantee service:

SONEL S.A.

Wokulskiego 11
58-100 Świdnica
Poland

tel. +48 74 858 38 60

fax +48 74 858 38 09

E-mail: export@sonel.pl

Web page: www.sonel.pl



ATTENTION!
Service repairs must be performed only by the manufacturer.



- AutoISO-2511 adapter is not a stand-alone measuring device. It is an accessory designed to cooperate with a dedicated meter (**MIC-2511**). The rules for using the adapter and performing measurements with it can be found in the meter's manual. Please get acquainted with it before proceeding with measurements.
- The manufacturer reserves the right to introduce changes in the design, accessories and technical data of the adapter.